



Dynamic Transducers and Systems 21592 Marilla St. • Chatsworth, CA 91311 • Phone 818-700-7818 www.dytran.com • e-mail: info@dytran.com

## **SPECIFICATIONS**

## MODEL 3032A MINIATURE QUARTZ SHEAR LIVM ACCELEROMETER

SPECIFICATION		VALUE			<b>UNITS</b>
PHYSICAL					
WEIGHT, LESS CABLE WEIGHT, INCLUDING CABLE SIZE (HEX x HEIGHT) MOUNTING PROVISION CONNECTOR, COAXIAL, MOUNTED AT END OF 18 IN CABLE CASE /CAP MATERIAL ELEMENT TYPE PERFORMANCE		1.5 5.6 .25 x .28 ADHESIVE MOUNT 10-32, UNF-2A TITANIUM QUARTZ SHEAR			GRAMS GRAMS INCHES JACK
SENSITIVITY, +/-10% [1] RANGE F.S. FOR +/- 5 VOLTS OUT FREQUENCY RESPONSE, +/- 10% FREQUENCY RESPONSE, +/- 3db MOUNTED RESONANT FREQUENCY, NOM. EQUIVALENT ELECTRICAL NOISE (RESOLUTION) AMPLITUDE NON-LINEARITY (ZERO BASED BEST FIT ST.LINE M TRANSVERSE SENSITIVITY, MAX. STRAIN SENSITIVITY	ETHOD)	10 +/- 500 1 to 10,000 .66 TO 12,000 80 .007 2.0 5	G's PER MICROS	F	mV/G G's Hz Hz kHz G, RMS F.S., MAX. PERCENT @ 250/μσ
ENVIRONMENTAL					
MAXIMUM VIBRATION MAXIMUM SHOCK TEMPERATURE RANGE THERMAL COEFFICIENT OF SENSITIVITY SEAL		1000 1500 -60 TO +250 0.06 EPOXY/WELDED			G's, RMS G's, PEAK <sup>O</sup> F %/ <sup>O</sup> F
ELECTRICAL					
EXCITATION (COMPLIANCE) VOLTAGE RANGE EXCITATION CURRENT RANGE OUTPUT IMPEDANCE, NOM. OUTPUT BIAS VOLTAGE DISCHARGE TIME CONSTANT, MIN. OUTPUT SIGNAL POLARITY FOR ACCELERATION TOWARD TOP	+20 to +	2 to 20 7.0 to 9.0 0.5 POSITIVE GOING		VDC OHMS	mA VDC SEC.

- [1] MEASURED AT 1 G RMS AT 100 HZ PER ISA RP 37.2
  [2] A CALIBRATION CERTIFICATE TRACEABLE TO **NIST** IS SUPPLIED WITH EACH INSTRUMENT.