

Dynamic Transducers and Systems 21592 Marilla St. • Chatsworth, CA 91311 • Phone 818-700-7818

www.dytran.com • e-mail: info@dytran.com

SPECIFICATIONS MODEL 3225E2 MINIATURE CHARGE MODE ACCELEROMETER

SPECIFICATION	VALUE	<u>UNITS</u>
PHYSICAL		
WEIGHT SIZE (DIA x LENGTH HEX x HEIGHT) MOUNTING PROVISION CONNECTOR, RADIALLY MOUNTED [1] CASE MATERIAL SENSING ELEMENT TYPE	0.6 0.31 x 0.46 x .19 FLAT MOUNTING SI 3-56 TITANIUM PIEZOCERAMIC PLA	GRAMS INCHES URFACE FOR ADHESIVE MOUNT JACK ANAR SHEAR
PERFORMANCE		
SENSITIVITY, ±20% [2] [3] FREQUENCY RESPONSE, ± 10 % MOUNTED RESONANT FREQUENCY, NOM. AMPLITUDE NON-LINEARITY (ZERO BASED BEST-FIT ST.LINE METHOD) TRANSVERSE SENSITIVITY, MAX. STRAIN SENSITIVITY	1.8 [3] to 10,000 40 2.0 5 .0005	pC/G Hz kHz % F.S., MAX. PERCENT G's PER MICROSTRAIN @ 250/μσ
ENVIRONMENTAL		
MAXIMUM VIBRATION MAXIMUM SHOCK TEMPERATURE RANGE THERMAL COEFFICIENT OF SENSITIVITY ENVIRONMENTAL SEAL	400 5000 -60 TO 350 SEE GRAPH THIS P HERMETIC	G's, RMS G's, PEAK °F
ELECTRICAL		

OUTPUT SIGNAL POLARITY FOR ACCELERATION TOWARD TOP CASE GROUNDING

GROUND ISOLATED

Positive


SUPPLIED ACCESSORIES: (1) Model 6591A Removal Tool.

(1) Model 6298 Small Petro Wax

NOTES:

- [1] CONNECTOR MATES ONLY WITH DYTRAN CABLE MODEL 6003AXX. (XX IS LENGTH IN FEET)
- [2] MEASURED AT 100 Hz, 1 G RMS PER ISA RP37.2.
- [3] ACTUAL SENSITIVITY IS GIVEN ON A CALIBRATION CERTIFICATE TRACEABLE TO NIST, SUPPLIED WITH EACH INSTRUMENT. LOW FREQUENCY RESPONSE IS DEPENDENT UPON THE DISCHARGE TIME CONSTANT OF THE CHARGE AMPLIFIER.

TYPICAL THERMAL SENSITIVITY GRAPH



OPERATING INSTRUCTIONS

REV B ECN 6935 08/30/10 2